

Declaration of Conformity TOPRO Troja 5G

Document ID:	6319	Document type:	Declaration of conformity
Revision:	3	Document owner:	Quality
Prepared date:	15.04.2025	Prepared by:	Cosmin Cioroiu
Last revised:	15.04.2025	Revised by:	Cosmin Cioroiu
Approved:	Yes	Approved by:	Nicolai Skarsgård
Approved date:	15.04.2025		

Published: 15.04.2025

This Declaration of Conformity is issued under the sole responsibility of the manufacturer.
This DoC is designed according to Annex IV of Medical Device Regulation (EU) 2017/745.

As Legal Manufacturer, we

TOPRO Industri AS•
Rambekkveien 1
NO-2816 Gjøvik
NORWAY

SRN: NO-MF-000003447

hereby declare under our sole responsibility that the following CE marked device(s)

Product/trade name(s)	TOPRO Troja 5G	
Intended Purpose	The device shall give support to users with reduced balance and/or reduced walking ability	
Product Number(s) and Descriptions	814600	TOPRO Troja 5G M
	814601	TOPRO Troja 5G S
Product Options/Accessories	113000 / 113010 / 113020 / 113030 / 113040 / 113050 / 113060 / 113070 / 113080 / 114000 / 114010 / 114020 / 114030 / 114040 / 114050 / 114060 / 114070 / 114080 / 113500 / 113570 / 113560 / 113510 / 114500 / 114570 / 114560 / 114510	
Basic UDI-DI	705432T5G8146NQ	

are classified per rule 1 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class I devices in accordance with all applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

Document title: **Declaration of Conformity TOPRO
Troja 5G**
Document ID: **6319**
Approved: **Yes**

Document type: **Declaration of conformity**
Revision: **3**
Approved by: **Nicolai Skarsgård**

The object of the Declaration described above is in conformity with the following regulations and standards:

EU Regulation	Medical Device Regulation (EU) 2017/745
NS-EN ISO 9999:2022	Assistive products - Classification and terminology
NS-EN ISO 21856:2022	Assistive products - General requirements and test methods
NS-EN 1985:1998	Walking aids - General requirements and test methods
NS-EN ISO 11199-2:2021	Assistive products for walking manipulated by both arms - Requirements and test methods – Part 2: Rollators
NS-EN ISO 20417:2021	Medical devices – Information to be supplied by the manufacturer

TOPRO hereby confirm that all models/variants and their original accessories are produced and tested in accordance the above mentions regulations and standards. All the technical documentation for the device(s) are stored at the manufacturer.

The user manuals are attached with the products.

GJØVIK/ 15.04.2025

Nicolai Skarsgård / CEO

Document history

Replacements

DATE	HISTORY	REV	SIGN
2021.05.04	Replaced ID 6043		ÅA

Changes

DATE	HISTORY	REV	SIGN
2023.02.06	Updated list of standards	2	LB
2025.04.15	Added new variants	3	SK